


APPLICATION RESULTS


COMPOUND IDENTIFICATION

Compound Name

trans-Stilbene Oxide



CHROMATOGRAM



ANALYTICAL REPORT

Column Trade Name	CHIRALPAK [®] IC-U		
Column Size (Particle Size)	3 x 100 mm (1.6 μ m)		
Mobile Phase Composition	n-hexane / 2-propanol = 90 / 10		
Chromatographic Mode	Normal Phase		

Flow rate (mL/min)	0.425	Temperature ($^{\circ}$ C)	25
Injection Volume (μ L)	0.00004 (mg)	Sample Concentration	
Detection	UV-VIS 230 nm		
Elution Order			

No.	Rt	K'	Alpha Value	Resolution
1	1.8	0.65	1 - 2	10.53
2	2.3	1.15		