

## APPLICATION RESULTS





## CHROMATOGRAM



## ANALYTICAL REPORT

Column Trade Name Column Size (Particle Size)		CROWNPAK <sup>®</sup> CR-I(+) 3 x 150 mm (5 μm)		
Mobile Phase Composition		CO2 / methanol / water / trifluoroacetic acid = 70 / 27 / 3 / 0.15		
Chromatographic Mode		SFC		
Flow rate (mL/min)	3.0		Temperature (℃)	30
Injection Volume (µL)	1		Sample Concentration	2.7 mg/mL
Detection	LC-MS			
Elution Order				
No. R	ī	К'	Alpha Va	lue Resolution
1 0.6			1 - 2	5.40
2 1.2	<u>)</u>			
omment				

This data was provided by Professor Bamba of Kyushu University.

## References ;

Ultrafast simultaneous chiral analysis of native amino acid enantiomers using supercritical fluid chromatography/tandem mass spectrometry

Journal of Chromatography A 1677 (2022) 463305