



APPLICATION RESULTS

COMPOUND IDENTIFICATION

Compound Name	Etomidate
---------------	-----------



CHROMATOGRAM



ANALYTICAL REPORT

Column Trade Name Column Size / Particle Size	CHIRALPAK® ID 250 x 4.6mm / 5µm
Mobile Phase Composition	n-hexane / ethanol / methanol / diethylamine = 90 / 5 / 5 / 0.1
Chromatographic Mode	Normal Phase Primary screening solvent (Immobilised CSP)

Flow rate (ml/min)	1	Injection amount (mg)	0.005
Temperature (°C)	25	Pressure (MPa)	
Detection	UV 240 nm	Elution order	

No.	Rt	K'	Alpha Value	Resolution
1	7.13	1.38	1 - 2	1.31
2	8.42	1.81		4